

DISTRICT BOARD OF TRUSTEES  
PENSACOLA JUNIOR COLLEGE  
1000 COLLEGE BOULEVARD  
PENSACOLA, FL 32504-8998

April 17, 2008

ADDENDUM NO. 1  
TO  
INVITATION TO BID (BID 7 - 2007/2008) for Dental Units  
FOR  
PENSACOLA JUNIOR COLLEGE

The following is in response to vendor questions received for the referenced BID:

1. Where are the Delivery Systems going to mount? Will it mount to the: Chair, Rear Cabinet, Side Cabinet, or Wall? Where will the Lights be mounted? Ceiling? Chair? Wall?

**Both the light and the delivery system will be mounted on the chair.**

2. Please explain item #3, adaptable for monitor cables.

**The light pole has to have the ability to house the monitor cables.**

3. Is the base swivel lock part of the light or chair?

**The base swivel lock is part of the chair.**

Angie C. Jones  
Director of Purchasing  
and Auxiliary Services

Please acknowledge and return this addendum with your proposal. If you have already submitted a bid, please return this addendum, with signed acknowledgment, by the proposal opening deadline, **April 23, 2008**.

ADDENDUM ACKNOWLEDGED BY:

\_\_\_\_\_

FIRM/ENTITY

\_\_\_\_\_

SIGNATURE OF REPRESENTATIVE:

\_\_\_\_\_